

## Item#6002

### GLOVE, GRAPE GRIP, NITRILE, PF EXAM, MD, 10/100

Perfect for use in medical facilities, these gloves are approved for use for medical exams. Follow safety regulations, and ensure the safety of your staff and patients by providing only medical grade gloves and equipment. These gloves are made from nitrile and are completely latex free. This is especially helpful in medical environments where patients may not be able to inform staff about their latex allergies or sensitivities. Containing no powder, they will not leave residue on clothing or skin.



<b>CASE PACK</b>	10/100	<b>MATERIAL</b>	NITRILE
<b>COUNTRY OF ORIGIN</b>	Malaysia	<b>COLOR</b>	PURPLE
<b>CASE DIMENSIONS L x W x H (in)</b>	13.4" X 9.8" X 9.8"	<b>CASE CUBE (ft)</b>	0.745
<b>CASE WEIGHT (lb)</b>	10.4	<b>NET WEIGHT (lb)</b>	9.8
<b>PALLET TI HI</b>	14 x 7	<b>UPC MASTER</b>	00811033010610
<b>UPC INNER</b>	811033010603	<b>UPC INDIVIDUAL</b>	-

# Item#6002



<b>GRADE</b>	Examination	<b>GLOVE POWDER</b>	Powder Free
<b>SIZE</b>	M	<b>LATEX-FREE</b>	Yes
<b>WEIGHT PER GLOVE (g)</b>	3.7	<b>WEIGHT STANDARD, M SIZE (g)</b>	3.7
<b>PALM WIDTH (mm)</b>	95	<b>LENGTH (mm)</b>	Min. 240
<b>CUFF THICKNESS (mm)</b>	Min. 0.05	<b>CUFF THICKNESS, ASTM (mm)</b>	N/A
<b>PALM THICKNESS (mm)</b>	Min. 0.06	<b>PALM THICKNESS, ASTM (mm)</b>	0.05
<b>FINGER THICKNESS (mm)</b>	Min. 0.07	<b>FINGER THICKNESS, ASTM (mm)</b>	0.05
<b>COMPLIANCE STANDARDS</b>	ASTM F1671, Chemo-Approved, Fentanyl Protection	<b>TEMPERATURE TOLERANCE (°C)</b>	180
<b>FEATURES</b>	Ambidextrous, Low Dermatitis, passed ASTM 1671		
<b>STORAGE CONDITIONS</b>	The gloves shall maintain their properties when stored in a dry condition. Avoid direct sunlight and excessive heat.		